



EMC Test Report Full Compliance Measurements



Megacon IsoBase (Isohub/IsoOut) Measurements performed 2011-09-27

Certificate of Test

Equipment Under Test

Manufacturer: Megacon AB
Type of product: Earth Fault Monitoring Unit
Model number: Isobase
Serial Number: 516111437
Tested on behalf of: Megacon AB

Summary

The equipment under test described in this report

complies with does not comply with

- Emission according to EN 61326-1:2006, Edition 2, class B, Electrical equipment for measurement, control and laboratory use.
- Immunity according to EN 61326-1:2006, Edition 2, Basic Requirements Electrical equipment for measurement, control and laboratory use.

Provided that modifications described in clause 5 are implemented.

The test results in this report only apply to the specified sample of EUT.

Exclusions

No exclusions were made from the standards.

Date: Nov 08, 2011

Tested by:



Michael Sohlberg

Approved by:



Johan Bergstrand



Note: This is an electronic copy. The original is a signed and stamped paper version.



Safety Test Report



Megacon IsoBase (Isobox16) – IsoHub/IsoOut

TEST REPORT	
EN 61 010-1 / IEC 61 010-1	
Safety requirements for electrical equipment for measurement, control and laboratory use	
Product:	Earth Fault Monitoring Equipment
Model / type:	IsoBase , Isobox16
Units included in the system:	IsoHub and IsoOut
Trade mark:	Megacon
Serial no:	516111437 (IsoBase)
Name and address of applicant:	Megacon AB Ranhammarsvägen 20 168 67 Bromma, Sweden
Name and address of factory:	Same as applicant
Standard:	EN 61010-1: 2010 3 rd edit IEC 61010-1: 2010 3 rd edit
The equipment complies with the standard:	PASS
National deviations:	EU
Other requirements:	-
Equipment mobility:	Fixed equipment
Operating condition:	Cont.
Class of equipment:	Class I
Protection against ingress of water:	IP20 verified in this report
Date: 2011-09-26 to 2011-10-19	
Tested by:	 Signature Michael Sohlberg
	BK Services AB Datalinjen 5A 583 30 Linköping, Sweden
Date: 2011-10-21	
Approved by:	 Signature Johan Bergstrand
	BK Services AB Datalinjen 5A 583 30 Linköping, Sweden
Tested at:	BK Services AB 583 30 Linköping Sweden